

Real-time Identification of EMG Signals and Hand Gesture Recognition using Accelerometer

Ramesh K ¹, Abishek T P ², Ajai M ³, Ajith G ⁴, Janarthanan M ⁵

Assistant Professor, Department of ECE, Adhiyamaan College of Engineering, Hosur, Tamilnadu, India ¹,

U.G Students, Department of ECE, Adhiyamaan College of Engineering, Hosur, Tamilnadu, India ^{2,3,4,5}

ABSTRACT: The solid framework is answerable for the development of human body; it likewise capacities to create movement and power. The investigation of the muscle movement assists with the advancement strategies and injury avoidance. Because of late advances in electromyographic (EMG) advances, muscle developments in sports would now be able to be obtained utilizing remote surface. The point of this examination is to explore information investigation utilizing EMG and furthermore the hand motion acknowledgment utilizing accelerometer. This examination has been actualized by leading a writing audit and a review of test contemplates. The overview of Biomechanists gave a knowledge into the sensor gadget used. The investigation of strong exercises during maximal dashing returned associated results. In The utilization of Independent Component Analysis (ICA) it was inspected that cross-talk can be decreased during sports developments. This examination applies ICA to EMG signals during isometric constrictions; little increment in connection was found now and again between the yield signals and in this manner the perfect signs.

KEYWORDS: ATmega 328, EMG sensor, Temperature, Accelerometer, Buzzer.

I. INTRODUCTION

Electromyography is an electro-indicative therapeutic strategy which will be utilized for assessing and recording the electrical action created by the solid framework. EMG are regularly performed utilizing an instrument called electromyography to supply a record called electromyogram.

The fundamental target of this work is to recognize the ongoing electromyographic signal examples. This system starts with the portrayal of the extraction of EMG signals following by the Recognition of hand motions. The acknowledgment of hand signals is completed utilizing the accelerometer sensor.

Here right now EMG sensor is utilized to recognize the solid exercises that are being seen by the anodes of the sensor that are in contact with the patients/competitors muscles, developments are watched, recorded and afterward sent to the medical clinic the executives if there should arise an occurrence of patients or to the mentor of the competitor, the information is sent by means of the zigbee transmitter that can be accessible for the administration and furthermore by means of the Bluetooth android application that can be helpful in the event of any crisis circumstances. An accelerometer sensor is additionally utilized in such manner to perceive the hand motions of the patients remarkably.

The EMG signal is a biomedical sign it has a principle highlight that can quantify electrical flows produced in muscles speaking to neuromuscular exercises. The investigations for the strategies for EMG signals are promptly accessible, the character and qualities of the sign are regularly appropriately comprehended and equipment executions are frequently made for different EMG signal related applications.

Organized by

Department of ECE, Adhiyamaan College of Engineering, Hosur, Tamilnadu, India

AI is utilized to take care of the issue of hand motion acknowledgment dependent on EMG signals. The most widely recognized classifiers for hand signal acknowledgment incorporate help vector machines, k-closest neighbors (k-NN), irregular timberland, direct discriminant investigation, counterfeit neural system (ANN), convolutional neural systems and gated repetitive unit organize.

The traditional highlights utilized for hand motion acknowledgment are characterized inside the accompanying areas: time space, similar to mean distinct amount (MAV) and 0 intersection (ZC); recurrence area, for example, middle recurrence and force range proportion; and time-recurrence space, for example, wavelets. Models bolstered these classifiers and have areas present high acknowledgment precision. Nonetheless, a large portion of them have to anticipate the finish of the motions and can't be applied during a continuous framework.

Additionally, a few analysts have researched the featureless methodology that legitimately sustains the pre-handled information without include extraction into a classifier, a procedure which can decrease the computational expense. Be that as it may, separating suitable highlights from handled information can fortify the inborn attributes of the EMG signal, of which the component determination is that the way to improve the segregation execution. In this way, hand motion acknowledgment stays an open subject for shiny new research.

Paper is sorted out as follows. Segment II depicts the related work or the current framework that had been completed by different scientists. We have rolled out some significant improvements to the current framework, portrayed quickly in the Section III. Segment IV presents trial results. At last, Section V presents end.

II. EXISTING SYSTEM

Electromyography (EMG) signals have been progressively utilized in the example acknowledgment framework and restoration in the previous decades. The works did by different specialists has found to make them slack in the reaction time and the precision levels. We have concocted an adjustment in the framework and acquainting the microcontroller ATmega 328 with control the absolute framework, EMG sensor to recognize the myographic signals from our solid framework and an accelerometer sensor to identify the hand motions.

III. PROPOSED SYSTEM


Identification of the EMG signals from the muscular system to the most accurate level can achieved by using the EMG sensor and the recognition of the hand gestures can be observed through the accelerometer sensor, below a block diagram is represented in such a way to identify the signals.

Organized by

Department of ECE, Adhiyamaan College of Engineering, Hosur, Tamilnadu, India

A. BLOCK DIAGRAM

Transmitter:



Receiver:




Fig 1: Block Diagram of Transmitter and Receiver

In the above block diagram microcontroller ATmega328 is used as an interfacing sensor, communication device, indication and display. We are using an accelerometer for sensing hand gestures, temperature sensor to find the body temperature, EMG sensor to analyse the muscular movements. All these sensors produce the analog voltage which is connected to the ADC pin of the ATmega328 microcontroller. The program is written in such a way to convert analog signal into meaningful real-time parameters. We are using a LCD display to monitor the human body parameters locally. We are extending the monitoring to remote locations using wireless communication devices like Bluetooth and Zigbee. The parameters are monitored on mobile using an android application and monitored using a PC (wireless application). There is a buzzer attached to the device to alert at any abnormality condition.

IV. EXPERIMENTAL RESULTS

Hand signals joined with arm stances were finished by certain subjects in the analyses. The subjects don't have any physical issues in making hand signals and are completely mindful of the test procedure. The subjects perform 10 preliminaries for single hand motion and rest for 1 min to stay away from the gathering of physical exhaustion. Every preliminary is executed in 1–2 s and the subjects have a rest for 5 s between preliminaries.

For 10 preliminaries, it takes around 1 min. In the trials, we educated subjects regarding the development timing by blaring. Therefore, in these tests, the beginning planning is known. The division is accomplished for 1 s after the beginning of blaring. Accordingly, the arrangement delay is the practically same as 1 s. The outcomes acquired from the analysis were arranged into cases that exclusively utilized EMG signals and those that all the while utilized EMG and accelerometer signals for correlation. The parameters N, M, and wmax for the waveform length and the histogram were set to 512, 50, and 200, separately. The hand motions were rehashed multiple times at every one of nine arm stances to figure the normal acknowledgment exactness. Twenty-five information focuses arbitrarily examined an aggregate of 50 preliminaries were utilized as the preparation set, and the staying 25 information focuses were utilized as the test set for cross-approval, which was rehashed multiple times to figure the normal acknowledgment precision. The lion's share casting a ballot has not been applied to the preparation model. The hand motions for F1, F2, and F3 were perceived as the arm pose changes from P4 to P7 and P8 successively and the perceived motions are set apart on the diagram. On account of the sole work of EMG flags, the hand motions F2 and F3 were mistakenly perceived as F4 and F5, separately. Nonetheless, the acknowledgment of hand motions that considered the comparing arm poses effectively separated all hand signals. So as to confirm the exhibition of the proposed technique, we contrasted it and the current classifiers, for example, straight, nonlinear SVM, and k-NN. The outcomes accomplished from the work will be transmitted through remote correspondence. Comparison of our proposed system and the other existing systems are represented in the table shown below.

Model	Accuracy %	Response(ms)
Proposed model	98.7%	227.76
Model using K-NN	89.5	246.2
Model using ANN	90.7	257.32
Model using Deep-Learning	98.31	269.12

Table I: Comparison of Accuracy and *Response*

V. CONCLUSION

This paper closes the ongoing ID of the EMG signs and acknowledgment of hand motions utilizing accelerometer. EMG sensor is utilized to gain EMG flags and applying a sliding window way to deal with fragment the information for extricating highlights. A feed-forward counterfeit neural system is established and prepared by the preparation dataset. A test technique is utilized that the motion will be perceived when perceived name times arrive at the edge of actuation times by the ANN classifier. The model reacts in 227.76 ms from the beginning of the motion, which is not exactly the cut off characterized for continuous (300 ms). Simultaneously, the model shows an acknowledgment precision of 98.7%, which is higher than the best in class.

REFERENCES

- [1] Ruize Xu, Shengli Zhou, and Wen J. Li, Fellow, IEEE, "MEMS Accelerometer Based Nonspecific-User Hand Gesture Recognition", IEEE SENSORS JOURNAL, VOL. 12, NO. 5, MAY 2012.
- [2] Prof. Vishal V. Pande, Nikita S. Ubale et.al "Hand Gesture Based Wheelchair Movement Control for Disabled Person Using MEMS.", IJERA, Vol. 4, Issue 4, April 2014.
- [3] Sakshi Goyal, Ishita Sharma, Shanu Sharma, "Sign Language Recognition System For Deaf And Dumb People", IJERT, Vol. 2 Issue 4, April - 2013.
- [4] Liu, S.H.; Cheng, W.C. Fall detection with the support vector machine during scripted and continuous unscripted activities. Sensors 2012, 12, 12301-12316.
- [5] Pendharkar, G.; Naik, G.R.; Acharyya, A.; Nguyen, H.T. Multiscale PCA to differentiate regular and irregular surfaces using tri axial head and trunk acceleration signals. In Proceedings of the 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBS), Milan, Italy, 25-29 August 2015; pp. 4122-4125
- [6] Xi, X.; Tang, M.; Miran, S.M.; Luo, Z. Evaluation of features extraction and recognition for activity monitoring and fall detection supported wearable EMG sensors. Sensors 2017, 17, E1229.

International Journal of Innovative Research in Science, Engineering and Technology
An ISO 3297: 2007 Certified Organization *Volume 9, Special Issue 1, March 2020*
8th National Conference on Frontiers in Communication and Signal Processing Systems (NCFCSPS '20)
18th-19th March 2020

Organized by

Department of ECE, Adhiyamaan College of Engineering, Hosur, Tamilnadu, India

- [7] He B., Wang S., Liu Y. Underactuated robotics: A review. *Int. J. Adv. Robot. Syst.* 2019;16:1–29. Zhai X., Jelfs B., Chan R.H., Tin C. Self-recalibrating surface EMG pattern recognition for neuroprosthesis control based on convolutional neural network. *Front. Neurosci.* 2017;11:379.
- [8] Zhang Y., Cao C., Cheng J., Lu H. Egogesture: a replacement dataset and benchmark for egocentric hand gesture recognition. *IEEE Trans. Multimedia.* 2018;20:1038–1050. doi: 10.1109/TMM.2018.2808769.
- [9] Taylor J., Luff B., Topalian A., Wood E., Valentin J. Efficient and precise interactive hand tracking through joint, continuous optimization of pose and correspondences. *ACM Trans. Graph.* 2016;35:1–12
- [10] Wu H., Guo H., Su Z., Shi M., Chen X., Cheng X., Han M., Zhang H. Fabric-based self-powered noncontact smart gloves for gesture recognition. *J. Mater. Chem.* 2018;6:20277–20288.